

Medicines Inspectorate Visit to B.P.L. including
Technical Services section on 5th/6th March 1981.

As a result of the above visit, A.T.S.M. received a
telephone call late on Friday evening from the Director,
Dr. R.S. Lane, in which it was intimated that the
operation of our moist steam sterilisers was poorly
managed, and that standards were inadequate.

He stated that he was considering the possibility
of shutting the operation down until the sterilisers
had been validated. He requested information on
thermocouple entry compression joints which would
enable us to start validation.

The T.S.M. advised that he hoped Vanley
Engineering would deliver the equipment on
Monday 9th March. Over the weekend,
the T.S.M. followed this point through and it
appears that the parts required will not be ready
until the following Monday, 16th March when
Mr Vanley will bring them to Chetree and
commission the equipment.

1248

20/273

Despite difficult circumstances, Technical Services Section has made considerable progress towards improved management of the sterilisers.

Routine daily/weekly operating checks and tests are now performed by the section and the engineers. (these were non-existent at time of the 1979 inspection).

- | | | |
|----|---|--------|
| 1 | Records and temperature gauge agreement | DAILY |
| 2 | Check for cleanliness | " |
| 3 | Air leak test (porous load type only) | WEEKLY |
| 4 | Bowie Dick test " " " " | DAILY |
| 5 | Temperature sensitive tape (all items) | " |
| 6 | Browne Control tubes " " | " |
| 7 | Door Seal check | WEEKLY |
| 8 | Door Safety devices | " |
| 9 | Steam traps | " |
| 10 | Clean chamber | " |

Periodic tests are performed by the Section and microbiology dept.

- | | | |
|----|---|---------------------------|
| 11 | Sponge strips tests (microbiological control) | |
| 12 | Air filters | down displacement MONTHLY |
| | porous load | 3 MONTHLY |
| 13 | Thermocouple tests to come on line during MARCH 1981. | |

20/274

The tests and checks are supervised for T.E. by Mrs. V. Plummer. Mr. Singleton is responsible for the growth and test results in microbiological test No. 11, while tests 3, 9 and 12 are performed by Mr. Montgomery's staff.

Validation of moist steam sterilisers.

Validation of standard loads is going to be very time consuming unless the sterilisers are essentially used for such tests, to the exclusion of routine production work. i.e. Taken out of normal service for a week each machine, and provided that staff can be available to ensure proper maintenance of equipment during this period. The sensible way to tackle this problem would be over a period of about 3 months during which time essential equipment be treated as priority for test.

Validation of sterilisers is usually the responsibility of two distinct groups of people.

Routine daily tests of basic nature are attended by the user i.e. operator and Supervisor. More complicated tests such as validation of standard loads and commission testing is NOT left to the user since this may be introduced i.e. the user may be tempted to establish test

20/275

conditions which are favourable to him.

In the NHS, thermocouple tests of sterilising equipment are routinely performed by the Hospital Engineers staff. In private industry this task is performed by the Q.C. Section.

During the past 10 years of operation, no commission tests have been performed, routine daily/weekly tests have been undocumented and mostly not performed. Documentation was non-existent, periodic tests were meaningless and results have never been reported to the T.S.M.

Wrapping of equipment.

Responsibility for wrapping has been divided between production departments and Technical Services, due to inadequate staff levels and space in T.S. area. During the visit by M.I., those items of equipment not properly wrapped, were mainly prepared by production units who complete their own assembly and wrap operation. Two main areas of complaint were noted:

- i Equipment from C.F. lab, particularly filters.
- ii Clothing for use in sterile areas.

A meeting of users will be arranged for week ending 13/3/81 to discuss correct procedure.

20/276

Problems in T.S. area when considering priorities of work:

1. Considerable time has been, and still is, being expended on management of sterilizers
2. Mrs Plummer is learning to supervise this type of service unit and is making progress. She was not conversant with the equipment or techniques prior to joining B.P.L. As a part of her training Mrs P. has visited Edgeware CSSD for instruction and advice.

Additionally there has been the need to share the T.S.M. office with Mrs P. This has been good from the viewpoint of regular meetings and induction, however it does cause increased use of the office by other staff and this leads to difficulty over continuity of work effort on my part.

3. Staff levels in T.S. are inadequate. A fully centralized service cannot be performed with present structure and it is unlikely that we could house sufficient staff to do so even after completion of MAP/01, let alone in present accommodation.
4. The T.S.M. has had several demanding responsibilities, all priority to deal with in addition to sterilizers.

20/277

- i Investigate, commission and organise S.O.P. operation of the new bottle washer.
- ii Take over and validate hot air sterilising oven.
Initiate documentation for bottles washed on machine and over sterilised, with batch identification.
Arrange scheduled cleaning of new machine.
STAFF LEVELS IN T.C. ARE INADEQUATE FOR THIS SERVICE
- iii Establish S.O.P for receipt of infected wastes and safe disposal of same.
PREVIOUSLY NO CONTROL OVER THESE PRODUCTS
- iv Meet nurse crosshairs cleaning managers, (3 in 7 months), during the formative first years service. A estimate that 50% of my time has been spent on documentation, control of contract and in particular trying to get BPL staff involved at all levels.
- v Clean room laundry & Finis Aera to organise shortly.

Equipment

Our sterilisers are inadequate by M.I. standards. BPL seems to have accepted that we have three porous load sterilisers because the manufacturers described them as such when supplied. The sterilisers do not conform to B.S.; in the case of the two largest machines it is incorrect to describe them as such, these machines are not suitable for wrapped instruments.

20/275

These two machines, the most useful as regards capacity, are best described as down displacement with a good vacuum assist and air drying.

The majority of the process equipment used in BPL process areas does not technically need to be sterile, and only needs to be clean and free from virus.

Wrapping in such circumstances helps to lessen the bacterial and particulate load added to the unstable product. - This may be desirable but is not essential to production.

The devices are therefore not ideal but in my opinion adequate for the basic task asked of them.

It should be stressed that we do not terminally sterilise any of our products.

Mr Vallet was asked if we should budget in 1981/82 for new B.S. sterilisers to the appropriate Paras Load Standard - however it was suggested that with the doubt about commercial involvement, the continued possibility of DHSS funding and perhaps plans for a new building, that it was not appropriate to do so.

20/279

Summary:

1. The inspectors were critical of our sterilisers, lack of thermocouple validation, wrapping techniques and of the congestion in T.S. area.
 - i. It was not considered appropriate to budget for new sterilisers when last discussed.
 - ii. Most of the equipment wrapping authorised by M.I. was prepared by C.F. Lab and Terminal Processing staff. We do advise on wrapping techniques but it is difficult to enforce standards unless under our direct control.
 - iii. Tests and checks of equipment have been increased, although validation with thermocouples remains to be introduced.

There can be little doubt that at the extremes of temperature/time exposure used by BPL, equipment will be sterilised.

Thermocouple tests will commence on 10/3/81 putting the wires via door seals. This preliminary work to be followed by tests via the proper entry ports on 16/3/81 when Valey Engineering commission the compression joints.

It should not be considered that T.S. underestimates the need for more accurate validation. The deficiencies of the system cannot however be tackled overnight bearing in mind deficiencies in staff numbers and experience.

20/25

11 It is not possible to employ sufficient numbers of staff within the T.S. area, thus ensuring a complete service, mainly due to lack of space. We have been criticised for congested work areas.

MAEP/01 will provide marginal relief with reference to congestion.

Additional staff levels have been discussed with administration although there is a need to agree grades of pay for new operatives. (see action list).

As soon as Mr Sharp (steriliser) retires, in May, it should prove possible to operate more frequent sterilising runs with smaller more controlled payloads. Restructuring of old established work patterns will lead to greater flexibility and will also ease congestion.

GRO-C: Mr Levens

20/251

Action List.

1. VARLEY ENGINEERING to commission extension joints 16.3.81

2. PSL To validate some loads with effect from 10.3.81 by use of Chesil equipment, putting thermocouples through chamber door seals.

Full validation work to commence 16.3.81

3. Arrange notes, meeting & demonstration of working techniques for CF lab and T.P. lab. (asap.)

4. Additional staff requirements.

Establishment.

STERILISERS - 2 OPERATIVES (in BPL)
FOR MOIST DRY HEAT
EQUIPMENT.
09.00 to 17.30.

EXISTING

STERILISERS / CLEAN LAUNDRY

- 2 OPERATIVES (in MEDIA)
09.00 to 16.00

ONE ADDITION
TO
EXISTING

All at ICS grade 0-2-4.

SUPERVISOR FOR STERILISERS / STILLS / OVENS
09.00 - 17.30
AT ICS grade 10

ONE
ADDITION
TO
EXISTING.

PROCESS OPERATIVES - 2 OPERATIVES
09.00 - 17.30
FOR SPECIAL RESPONSIBILITIES
EX CF, T.P. AREA.
WASH/ASSEMBLY EQUIPMENT
WASHING MACHINE OPERATION
PLASMA DISPOSAL
(GRADE TO BE AGREED)

TWO
ADDITIONS
TO
EXISTING

20/252