

Bled 9/7/86  
Tested 10/7/86

\*\*\*\*\*  
HTLV 3 SCREEN  
\*\*\*\*\*

Disc : WELLCOME / HTLV3 DATA DISC 3

PLATE NUMBER : 10/24

DATED 10/07/86 at 11:51

CUTOFF= 691 MR600 CALIBRATION= 1000 TEST START No. 1 at well A 1

Controls

Strong Positive/s at : H 8 ,  
Weak Positive/s at : E 8 , F 8 , G 8 ,  
Negatives at : C 8 , D 8 ,

1 2 3 4 5 6 7 8 9 10 11 12

A:	947	1403	1502	1879	1484	1493	1472	1537			
B:	1143	1173	1205	1246	1248	1246	1201	1095			
C:	1452	1317	1384	1420	1363	1393	1273	1400			
D:	1436	1376	1315	1381	1344	1284	1299	1418			
E:	1490	1438	1304	1336	1410	1274	1361	630			
F:	1437	1384	1346	1368	1354	1288	1320	637			
G:	1493	1409	1395	1358	1407	1391	1392	619			
H:	1540	1378	1324	1389	1342	1276	1353	115			

A:	-	-	-	-	-	-	-	-			
B:	-	-	-	-	-	-	-	-			
C:	-	-	-	-	-	-	-	-			
D:	-	-	-	-	-	-	-	-			
E:	-	-	-	-	-	-	-	-	+		
F:	-	-	-	-	-	-	-	-	+		
G:	-	-	-	-	-	-	-	-	+		
H:	-	-	-	-	-	-	-	-	+		

3293729-779(1-51) P307591-595(52-56) MIS E292181(57) E292189(58)  
E292197(59) (NEG CONTROL 1)

ALL TESTS NEGATIVE

ALL TESTS NEGATIVE

TOTAL TESTS = 59

POSITIVES = 0  
% POSITIVE = 0.0


QUERIES = 0

Read by : \*\*\*\*\*Checked by : \*\*\*\*\*

No 25 .6293.753.

## DISTRIBUTION GRAPH

PLATE NO. 10/24



CUTOFF= 691

NEGATIVE RESULTS -

MEAN= 1359.793

STANDARD DEVIATION= 127.2

RANGE from 947 to 1879